

DEFIBRILLATOR STATUS REPORT

September 2026.

There have been no concerns since the previous update, (May 2026).

The regular weekly inspections have confirmed that the indicators have shown full function of all modalities – electronic circuitry , pads, and battery.

The machine continues to be logged with “The Circuit” (National Defibrillator Network) each month.

Replacement of “consumables” is not due until 2028 unless the indicators show otherwise.

Another individual will be needed to continue the surveillance of the machine in the not too distant future. I will give good notice of when this will be necessary. If there is anyone who would like to volunteer for this rôle , then I’ll go through everything with them and log them in as a “Circuit Guardian”. Until then I will continue to fulfil this rôle, if everyone is happy with this.

Wendy Buswell

Circuit Guardian.

23/09/26